Setting Up the EPP

1 Press the PAIRING button for 5 seconds.

2 Move the EPP at least 50 cm (20 in.) away from the Control Panel.

50 cm (20 in.)

3 Press the button.

The panel beeps and the EMERGENCY button blinks blue.

EMERGENCY CALL

Blue

The EPP in Action

Panic Button

Monitoring Station

The EPP blinks **red** when the panic button is pressed or a fall is detected

The EPP lights **green** when the battery is replaced

For more information, see:

www.adtsecurity.com.au/home-assure

Phone 131 005

ADT **Home Assure**

ADT **Home Assure**

PERS

Home Assure
Control Panel and EPP Kit

Getting Started

RESET

ADT

S.O.S

The Best Location for the Control Panel

Good cellular reception

Convenient for resident to access

Where speakerphone can be heard throughout premises

**1** Place thumb in oval hold and lift off the cover

2 Check SIM is inserted or insert

Check backup battery is inserted or insert

3

4

Connect the power cable and plug into a power source. Then close the back cover

5 Turn the power switch on

OFF | ON

RESET

Press to cancel an emergency call

Press to start an emergency call. Red light:

- Blinking: alarm in progress
- On: message acknowledged

PAIRING button

Shows power status

- On: AC power connected
- Blinking: using battery
- Off: no power

Shows communication status

- **Green**: if on, ready for call. If blinking, message being sent
- **Red**: if on, communication failure
- Orange: active conversation

Shows problem status

- On: malfunction in the Control Panel or a peripheral
- Blinking: supervision lost, or low battery in a peripheral